

QA File Parameter

Patient Name

:

Patient ID

:

WP-001

Plan Date

:

2020-09-25

SSD

:

100.0

SDD

:

Depth

:

Energy

:

6x

Angle

:

Absolute Dose Comparison

Difference (%)

:

2

Distance (mm)

:

2

Threshold (%)

:

10.0

Meas Uncertainty

:

No

Use Global %

:

Yes

Summary (Gamma Analysis)

Total Points

:

68

Passed

:

68

Failed

:

0

% Passed

:

100

Dose Values in cGy

	CAX	Norm	Sel	Max
MapCHECK2	185.79	184.12	185.79	186.52
Plan	186.60	186.50	186.60	186.60
MapCHECK2	-0.81	-2.38	-0.81	-0.08
-Plan				
% Diff	-0.44	-1.27	-0.44	-0.04
DTA(mm)	0.00	0.00	NA	
Coords (y,x) cm	0,0	-1,0	0,0	

Notes

2% 2mm TB4

Reviewed By :

MapCHECK2

...\\5by5 6x.txt

MapCHECK2-Plan

Plan

CAX Offset X=-2 Y=1

...\\Water_...